

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

JAN 21 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130176</u>
Name <u>DARYL RICE</u>		ID# <u>2088</u> Date <u>01/13/10</u>
A	Agency <u>SEIKA POLICE DEPT.</u>	Phone # <u>747-3245</u>
Instrument Location <u>304 LAKE STREET</u>		
Alco S/N <u>X301572</u> Target Value <u>.078</u> High Pressure <u>1200</u>		
B	Alco Test Values <u>.080</u> <u>.080</u> 1 st Alco 2 nd Alco	
Signature <u>Daryl Rice</u>		<u>CB</u> <u>1/22/10</u>
(OVER)		

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (to be filled in this area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this _____ day of _____, 2010.

(Notary Seal Stamp)

Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130176

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

JANUARY 13, 2010

HC PSE 1200
01/13/10

OPERATOR'S NAME:

RICE DARYL DEAN

OPERATOR'S NUMBER: 2088

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

VOC

D.L. #: VOC

DEPT/AGENCY: SIT1

CASE/REPORT: VOC

TEST TYPE: U

ALCO TARGET VALUE: .078

ALCO S/N: X301572

--- BREATH ANALYSIS ---

.078 ADJUSTED FOR 29.73 in
ALCO TARGET .077 06:56
BLANK TEST .000 06:57
INTERNAL STANDARD (VERIFIED) 06:57
ALCO TU 29.73 in .000 06:57
BLANK TEST .000 06:58
SUBJECT SAMPLE (VBS 112210) .000 06:59
BLANK TEST .000 06:59
ALCO TU 29.73 in .000 07:00
BLANK TEST .000 07:00

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130176

JANUARY 13, 2010

TIME 06:54

HC PSE
1200

01/13/10

--- DIAGNOSTIC CHECK ---

COMPUTER:

OKAY

PROGRAM:

OKAY

SOFTWARE DATE:

01/08/09

HEATERS

SAMPLE CHAMBER:

49c

BREATH TUBE:

41c

BAROMETER:

29.73 in

FLOW DETECTOR:

OKAY

PUMP

HIGH SPEED:

OKAY

DETECTOR:

OKAY

FILTERS:

OKAY

QUARTZ STANDARD:

OKAY

CALIBRATION:

OKAY

PRINTER TEST

!\"#\$%&'()*+,-./0123456789:;<=>?@AB CDEFG
HIJKLMN OPQRSTU VWXYZ[\]^_`abcde fghijklmno
pqrstuvwxyz{|}~■